

EU DECLARATION OF CONFORMITY

DOC No.	DOC-MYMEDIKAL-SRT-004		
EC Certificate	Not applicable (Self- declared)		
Manufacturer	MY TICARET VE MEDIKAL A.S.		
Manufacturer Address	Ömerli mah General Şükrü Koraltı Cd no:33, 34555 Arnavutkoy/Istanbul, Turkey		
Single Registration Number (SRN)	TR-MF-000018372		
Brand	B-good		
Product Description	Powdered Latex Examination Gloves		
Intended Purpose	A patient examination glove is a medical device intended for a medical purpose that is worn on the examiner’s hand or finger to prevent contamination between the patient and examiner. Examination glove is intended for medical activities except for surgery.		
Basic UDI-DI	868227994LP5H		
Size	XS, S, M, L, XL		
European Medical Device Nomenclature (EMDN)	T010201 (Examination/Treatment Gloves, Latex)		
Global Medical Device Nomenclature (GMDN)	47173 (Latex examination/treatment glove, powdered)		
Product Catalogue/Reference Number	BGL01-XS, BGL02-S, BGL03-M, BGL04-L, BGL05-XL		
Product Group Reference Number	LX01		
Conformity Assessment Route (MDR):	Annex II and Annex III according to EU 2017/745		
Classification & Rule (MDR)	Class I, Rule 5 transient use according to Annex VIII		
Device Classification (PPER)	Category III		
EU Type-Examination Certificate (PPER)	2777/10468-05/E05-01		
Notified Body (PPER)	EU-Type Examination and Ongoing Conformity by Notified Body SATRA TECHNOLOGY EUROPE LTD Bracetown Business Park, Clonee, D15YN2P, Ireland [CE 2777]		
Applicable Standards			
	No.	Regulation/ Standard Number	Regulation/ Standard Name
	1	MDR (EU) 2017/745	Medical Device Regulation



MY Medikal

MY TICARET VE MEDİKAL A.Ş.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –İstanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

	2	PPE (EU) 2016/425	Personal Protective Equipment Regulation
	3	ISO 13485: 2016	Medical devices - Quality management systems - Requirements for regulatory purposes
	4	ISO 9001: 2015	Quality management systems – requirements
	5	ISO 14971: 2019	Medical devices - application of risk management to medical devices
	6	EN 455-1: 2020	Requirements and testing for freedom from holes
	7	EN 455-2: 2015	Requirements and testing for physical properties
	8	EN 455-3: 2015	Requirements and testing for biological evaluation
	9	EN 455-4: 2009	Requirements and testing for shelf-life determination
	10	ISO 10993-10: 2010	Biological evaluation of medical devices –Part 10: Test for irritation and skin sensitization
	11	ISO 20417:2021	Medical devices - Information to be supplied by the manufacturer
	12	ISO 15223-1: 2021	ISO 15223-1 Symbols to be used with information to be supplied by the manufacturer
	13	EN ISO 374-1: 2016	Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks
	14	EN ISO 374-2:2014	Protective gloves against dangerous chemicals and micro-organisms - Part 2: Determination of resistance to penetration
	15	EN ISO 374-4:2013	Protective gloves against chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals
	16	EN ISO 374-5: 2016	Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risks
	17	EN 16523-1: 2015	Determination of material resistance to permeation by chemicals - Part 1: Permeation by



MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

		liquid chemical under conditions of continuous contact
--	--	--

We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:

- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentation is retained under the premise of the manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System.
- Is following the EU-Type Examination with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.
- Is in conformity to type based on quality assurance of the production process under the surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.
- This EU Declaration of Conformity is prepared in accordance with Annex IV of Medical Device Regulation (EU) 2017/745.

Authorized Signatory:

Approver : MURAT YILDIZ
Title : General Manager/CEO
Signature :
Approval Date : 15 Feb 2024
Place of Approval : Istanbul, Turkey

MY TICARET VE
MEDİKAL ANONİM ŞİRKETİ
Ömerli Mah. General Şükrü Koraltı Cad
No:33 Arnavutköy/İSTANBUL
Büyükdere V.D. 826 040 4605
Tel: 0212 438 20 64 Fax: 0212 438 20 65
www.mymedikal.com

