



As of 04 Mar 2024



USER INFORMATION

CE 2777

Comply with PPE Regulation 2016/425 Cat III

These products are classed as Category III Personal Protective Equipment (PPE) by the European PPE REGULATION 2016/425 and have been shown to comply with this Regulation through the Harmonised European Standard(s): EN ISO 21420:2020, EN ISO 374-1:2016+A1:2018, EN ISO 374-5:2016

Product references:

Mumu, Douromed, Elastron Latex Powderfree Gloves, Chlorinated, Non-sterile
ML01-XS, ML02-S, ML03-M, ML04-L, ML05-XL

B-good Latex Powderfree Gloves, Chlorinated, Non-sterile
BGLP01-XS, BGLP02-S, BGLP03-M, BGLP04-L, BGLP05-XL

Sizes available: XS, S, M, L, XL

Colour: Natural/ Pale Yellow, the color may vary due to storage time and conditions

Intended Use:

A patient examination glove is a medical device intended for a medical purpose that is worn on the examiner's hand or finger to prevent contamination between the patient and examiner. Examination glove is intended for medical activities except for surgery.

Performance and limitation of use –This product has been tested and achieved the following performance levels:

Classification:

EN ISO 374-1:2016+A1:2018 /Type B	Level	EN ISO 374-4:2019 Degradation%
40% Sodium Hydroxide (K)	4	-72.4%
96% Sulphuric Acid (L)	1	92.9%
65% Nitric Acid (M)	1	20.9%
37% Formaldehyde (T)	6	-81.7%
30% Hydrogen Peroxide (P)	6	-64.1%

EN ISO 374-5:2016

Protection against Bacteria and Fungi	Pass
Protection against Viruses	Pass

EN ISO 374-1:2016+A1:2018 /Type B



KTP



Virus





EN ISO 374-1:2016+A1:2018 Permeation levels are based on breakthrough times as follows:

Permeation performance level	1	2	3	4	5	6
Measured breakthrough time (min)	> 10	> 30	> 60	> 120	> 240	> 480

EN ISO 374-4:2019 Degradation results indicate the change in puncture resistance of the gloves after exposure to the challenge chemical.

EN ISO 374-5:2016 The penetration resistance has been assessed under laboratory conditions and relates only to the tested specimen.

This information does not reflect the actual duration of protection in the workplace and the differentiation between mixtures and pure chemicals.

“The chemical resistance has been assessed under laboratory conditions from samples taken from the palm only (except in cases where the glove is equal to or over 400 mm - where the cuff is tested also) and relates only to the chemical tested. It can be different if the chemical is used in a mixture.”

“It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on temperature, abrasion and degradation.”

“When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves”

“Before usage, inspect the gloves for any defect or imperfections.”

Storage and transport: When not in use, store the product in a well-ventilated area away from extremes of temperature

Glove performance quoted is based on laboratory data and may not reflect the actual duration of protection in the workplace due to other factors influencing the performance such as temperature, abrasion, degradation etc.)

The glove does not contain any substances that are known to cause allergies.

The Gloves have no mechanical protection offered.

For single use only, do not littering.

Check for damage before use, do not use damaged gloves





Donning:

1. Remove all hand and wrist jewelry, and wash the hands before donning.
2. Place the gloves on the prepared work surface.
3. The user puts a glove on his/her dominant hand by grabbing it with the other hand, remembering to only touch the inside of the gloves, and slipping it over the dominant hand until it reaches the final level.
4. The wearer uses the gloved dominant hand to slip the other glove onto the non-dominant hand.
5. Once both gloves are on, the users can touch the outside of the gloves to ensure a proper fit

Doffing:

1. Using the dominant hand, users start by grabbing the outside of the glove on the non-dominant hand on the palm side near the cuff.
2. Pull the glove off the non-dominant hand and place it in the gloved hand, balling it up.
3. Slip two fingers under the cuff of the other hand glove and carefully peel it off the hand without touching the wrist, turning the remaining glove inside out as it is removed and in turn encasing the first glove.
4. The gloves can be disposed.

The DOC (declaration of conformity) will be shown on website: www.mymedikal.com.tr/documents

Notified Body responsible for certification and ongoing conformity:

SATRA Technology Europe Ltd

Bracetown Business Park

Clonee, Dublin

D15 YN2P, Ireland (NB2777)

Product manufactured by: MY TICARET VE MEDİKAL A.Ş.





These products also comply with Medical Device Regulation (EU) 2017/745 under the Class I category.

CAUTIONS AND RESTRICTION

1. Keep in dry conditions and avoid sunlight or other ozone sources.
2. For single-use only.
3. For latex gloves: This product contains natural latex which may cause allergic reactions.

DISPOSAL METHOD

Throw it away in the special rubbish can after using it.

INSTRUCTIONS FOR USE

1. Always wash hands before putting on gloves and each time you change to a new pair.
2. Take the gloves out from their original box.
3. Select a glove appropriate for the task and in your size. Inform yourself prior to the application.
4. Touch only a restricted surface of the glove corresponding to the wrist (at the top edge of the cuff)
5. Don the first glove by inserting your hand into the opening and pulling it over by applying a soft pull with the second hand.
6. Take the second glove with a bare hand and touch only a restricted surface of the glove corresponding to the wrist.
7. To avoid touching the skin of the forearm with the gloved hand, turn the external surface of the glove to be donned on the folded fingers of the gloved hand so there is a good hold on it. Thus, permitting to glove the second hand by pulling softly with the first gloved hand.
8. Once gloved, hands should not touch anything else that is not defined by indications and conditions for glove use.
9. Change gloves before beginning a different task to avoid cross-contamination.
10. Pinch one glove at the wrist level to remove it, without touching the skin of the forearm, and peel it away from the hand, thus, allowing the glove to turn inside out.
11. Hold the removed glove in the gloved hand and slide the fingers of the ungloved hand inside between the glove and the wrist of the gloved hand. Remove the second glove by rolling it down the hand and fold the first glove into the outside palm area of the second glove. If done correctly, the user holds a pack of the outer glove turned inside and out and the contaminated surface of both gloves is within the package.
12. Discharge the removed gloves and wash hands thoroughly according to the respective hygiene standards.
13. Not intended to be used as a chemical barrier. The shelf life is normally advised to be





five years. Check the original dispenser instructions for details.

Other Common Graphical Symbols:

SYMBOLS	TITLE
	Manufacturer
	Date of Manufacture
	Use-by date/ Expiration Date
	Batch Code/ Lot Number
	Keep away from sunlight
	Keep dry
	Do not re-use
	Non-sterile

**MY TİCARET VE
MEDİKAL ANONİM ŞİRKETİ**
Ömerli: Mh.General Şükrü Koralı Cad
No: 33 Arnavutköy/İSTANBUL
B.ÇEKMECE V.D.6260404505
İl:0212 436 20 64 Fax:0212 436 20 65
www.mymedikal.com

