

EU DECLARATION OF CONFORMITY

DOC No.	DOC-MYMEDİKAL-ITC-004		
EC Certificate	Not applicable (Self- declared)		
Manufacturer	MY TICARET VE MEDİKAL A.S.		
Manufacturer Address	Ömerli mah General Şükrü Koraltı Cd no:33, 34555 Arnavutkoy/Istanbul, Turkey		
Single Registration Number (SRN)	TR-MF-000018372		
Brand	Mumu Guard DV		
Product Description	Nitrile Powder Free Examination and Protective Gloves		
Intended Purpose	A patient examination glove is a medical device intended for a medical purpose that is worn on the examiners hand or finger to prevent contamination between patient and examiner. Examination glove is intended for medical activities except surgery.		
Basic UDI-DI	868302002NPVQ		
Size	XS, S, M, L, XL		
European Medical Device Nomenclature (EMDN)	T01020204 (Examination / Treatment Gloves, Nitrile)		
Global Medical Device Nomenclature (GMDN)	56286 (Nitrile Examination/Treatment glove, non-powdered, non-sterile)		
Product Catalogue/Reference Number	MGDV01-XS, MGDV02-S, MGDV03-M, MGDV04-L, MGDV05-XL		
Product Group Reference Number	SNBE20013, SNBE20014, SNBE20015, SNBE20016, SNBE20017		
Conformity Assessment Route (MDR):	Annex II and Annex III according to EU 2017/745		
Classification & Rule (MDR)	Class I, Rule 1 & Rule 5 according to Annex VIII		
Device Classification (PPER)	Category III		
EU Type-Examination Certificate (PPER)	2777/14815-03/E00-00		
Notified Body (PPER)	EU-Type Examination and Ongoing Conformity by Notified Body SATRA TECHNOLOGY EUROPE LTD Bracetown Business Park, Clonee, D15YN2P, Ireland [CE 2777]		
Applicable Standards			
	No.	Regulation/ Standard Number	Regulation/ Standard Name
	1	MDR (EU) 2017/745	Medical Device Regulation
	2	PPE (EU) 2016/425	Personal Protective Equipment Regulation
	3	ISO 13485: 2016	Medical devices - Quality management systems -



MY Medikal

MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –İstanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

		Requirements for regulatory purposes
4	ISO 9001: 2015	Quality management systems – requirements
5	ISO 14971: 2019	Medical devices - application of risk management to medical devices
6	EN 455-1: 2020	Requirements and testing for freedom from holes
7	EN 455-2: 2015	Requirements and testing for physical properties
8	EN 455-3: 2015	Requirements and testing for biological evaluation
9	EN 455-4: 2009	Requirements and testing for shelf-life determination
10	ISO 10993-1: 2018	Biological evaluation of medical devices –Part 1: Evaluation and testing within a risk management process
11	ISO 10993-10: 2010	Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization
12	ISO 10993-11: 2017	Biological evaluation of medical devices — Part 11: Tests for systemic toxicity
13	ISO 20417:2021	Medical Devices- Information to be supplied by the manufacturer
14	ISO 15223-1: 2021	Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements
15	EN ISO 374-1: 2016+A1: 2018	Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks
16	EN ISO 374-2: 2019	Protective gloves against dangerous chemicals and micro-organisms - Part 2: Determination of resistance to penetration
17	EN ISO 374-4: 2019	Protective gloves against chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals
18	EN ISO 374-5: 2016	Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risks

	19	EN 16523-1: 2015+A1: 2018	Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact
	20	ASTM D 6978-05:2019	Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs
	21	ASTMF1671/F1671-13	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens Using Phi-X174 Bacteriophage Penetration as a Test System

We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:

- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentations are retained under the premise of manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System.
- Is following to the EU-Type Examination with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.
- Is in conformity to type based on quality assurance of the production process under surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.
- This EU Declaration of Conformity is prepared in accordance to Annex IV of Medical Device Regulation (EU) 2017/745.

Authorized Signatory:

Approver : MURAT YILDIZ

Title : General Manager/CEO

Signature

Approval Date

Place of Approval

MY TICARET VE
MEDİKAL ANONİM ŞİRKETİ
Ömerli Mah. General Şükrü Koraltı Cad
No:33 Arnavutköy/İSTANBUL
Büyükdere V.D.626 040 4605
Tel:0212 438 20 64 Fax:0212 438 20 65
www.mymedikal.com

:15 Feb 2024

: Istanbul, Turkey

CE