



As of 05 Aug 2024



USER INFORMATION

CE 2777

Comply with PPE Regulation 2016/425 Cat III

These products are classed as Category III Personal Protective Equipment (PPE) by the European PPE REGULATION 2016/425 and have been shown to comply with this Regulation through the Harmonized European Standard(s): EN ISO 21420:2020, EN ISO 374-1:2016+A1:2018, EN ISO 374-5:2016.

Nitrile Powderfree Examination and Protective Gloves

Brand	Catalog No./Ref. No./ Models
Mumu Plus	MN01

Sizes available: XS, S, M, L, XL

Color: Black

Intended Use:

A patient examination glove is a medical device intended for a medical purpose that is worn on the examiner's hand or finger to prevent contamination between the patient and examiner. Examination glove is intended for medical activities except for surgery.

Performance and limitation of use –This product has been tested and achieved the following performance levels:

PPE SPECIFICATION	PERFORMANCE LEVELS
Dexterity	5
Material Resistance To Permeation By Chemicals	2 (Type C)
Phi-X174 Bacteriophage	Appropriate

OBTAINED VALUE: EN 374-1/Type C, EN ISO 374-5

Protection against virus:

Evaluation	Limit Value
Results	Appropriate





Determination of material resistance to permeation by chemicals (%40 NaOH):

Evaluation	Limit Value
PERFORMANCE LEVEL:2	> 30 min

The permeation performance against Sodium hydroxide 40%

Evaluation	Limit Value
PERFORMANCE LEVEL:2	> 30 min

Recommendations for use:

- The PPE is Made from Synthetic Nitrile Rubber
- Powder-free
- Not made with natural rubber latex
- Ambidextrous

Warnings:

- Gloves not valid in an explosive atmosphere.
- This product is not flame resistant and must not be used in areas with open flames.
- The user must know the use and gloves handling.
- Do not use risk zones of explosion.
- This PPE must not be used against risks other than those previously described
- For single-use only.

Before usage, inspect the gloves for any defects or imperfections.

Storage and transport: When not in use, store the product in a well-ventilated area away from extremes of temperature

Check for damage before use, do not use damaged gloves.



**Donning:**

1. Remove all hand and wrist jewelry, and wash the hands before donning.
2. Place the gloves on the prepared work surface.
3. The user puts a glove on his/her dominant hand by grabbing it with the other hand, remembering to only touch the inside of the gloves, and slipping it over the dominant hand until it reaches the final level.
4. The wearer uses the gloved dominant hand to slip the other glove onto the non-dominant hand.
5. Once both gloves are on, the users can touch the outside of the gloves to ensure a proper fit

Doffing:

1. Using the dominant hand, users start by grabbing the outside of the glove on the non-dominant hand on the palm side near the cuff.
2. Pull the glove off the non-dominant hand and place it in the gloved hand, balling it up.
3. Slip two fingers under the cuff of the other hand glove and carefully peel it off the hand without touching the wrist, turning the remaining glove inside out as it is removed and in turn encasing the first glove.
4. The gloves can be disposed.

The DOC (Declaration of Conformity) will be shown on the website: www.mymedikal.com.tr

Notified Bodies responsible for certification and ongoing conformity:

EU-Type Examination by MNA LABORATUVARLARI SAN. TİC. LTD. ŞTİ. Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Ataşehir/ İstanbul, Turkey [Notified Body No.2841]	Ongoing Conformity by Notified Body SATRA TECHNOLOGY EUROPE LTD Bracetown Business Park, Clonee, D15YN2P, Ireland [CE 2777]
--	---

Product manufactured by: MY TICARET VE MEDİKAL A.Ş.





These products also comply with Medical Device Regulation (EU) 2017/745 under the Class I category.

Other Common Graphical Symbols:

SYMBOLS	TITLE
	Manufacturer
	Date of Manufacture
	Use-by date/ Expiration Date
	Batch Code/ Lot Number
	Keep away from sunlight
	Keep dry
	Do not re-use
	Non-sterile

MY TİCARET VE
MEDİKAL ANONİM ŞİRKETİ
Ömerli Mah. General Şükrü Koralı Cad.
No: 33 Arnavutköy/İSTANBUL
B.ÇEKMECE V.D.6260404505
İl:0212 436 20 64 Fax:0212 436 20 65
www.mymedikal.com

