

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

EU DECLARATION OF CONFORMITY

DOC No.	DOC-MYMEDIKAL-ZBS-001			
EC Certificate	Not applicable (Self- declared)			
Manufacturer	MY TICARET VE MEDIKAL A.S.			
Manufacturer Address	Ömerli mah General Şükrü Koraltı Cd no:33, 34555			
	Arnavutkoy/Istanbul, Turkey			
Single Registration Number	TR-MF-000018372			
(SRN)				
Brand	Mumu Guard			
Product Description	Nitrile Powder Free Examination and Protective Gloves			
Intended Purpose	A patient examination glove is a medical device intended for a medical purpose that is worn on the examiner's hand or finger to prevent contamination between the patient and examiner. Examination glove is intended for medical activities except for surgery.			
Basic UDI-DI	868302002NPVQ			
Size	XS, S, M, L, XL			
EAN Codes	Blue: 8684266526195, 8684266526201, 8684266526218, 8684266526225, 8684266526232 8683020024878, 8683020024885, 8683020024892, 8683020024908, 8683020024915 Black: 8684266526317, 8684266526324, 8684266526331, 8684266526348, 8684266526355 8684266521374, 8684266521381, 8684266521398, 8684266521404, 8684266521411			
European Medical Device Nomenclature (EMDN)	T01020204 (Examination / Treatment Gloves, Nitrile)			
Global Medical Device Nomenclature (GMDN)	56286 (Nitrile Examination/Treatment glove, non-powdered, non-sterile)			
Product Catalogue/ Reference Number	Blue: MGBSNPF01-XS, MGBSNPF02-S, MGBSNPF03-M, MGBSNPF04-L, MGBSNPF05-XL			



Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

	Black: MGBBSNPF01-XS, MGBBSNPF02-S, MGBBSNPF03-M,				
	MGBBSNPF04-L, MGBBSNPF05-XL				
Product Group Reference	Blue:				
Number	BS0102016, BS0102017, BS0102018, BS0102019,				
	BS0102020				
	Black:				
	BS0102056, BS0102057, BS0102058, BS0102059, BS0102060				
Conformity Assessment Route	Article 52(7) and				
(MDR):	Annex VIII, 4.1 Rule 1, Non-invasive, and/or				
	5.1 Intended for transient use, Rule 5 of invasive device				
Classification & Rule (MDR)	Class I, Rule 1 & Rule 5 according to Annex VIII				
Device Classification (PPER)	Category III				
EU Type-Examination	2777/21024-02/E13-02				
Certificate (PPER)					
Notified Body (PPER)	EU-Type Examination and Ongoing Conformity				
	by Notified Body SATRA TECHNOLOGY EUROPE LTD				
	Bracetown Business Park, Clonee, D15YN2P, Ireland [CE 2777]				
	Cionec, DISTINZI , II ciana [CL 2777]				
Applicable Standards					
	No.	Regulation/ Standard Number	Regulation/ Standard Name		
	1	MDR (EU) 2017/745	Medical Device Regulation		
	2	PPE (EU) 2016/425	Personal Protective Equipment Regulation		
	3	ISO 13485: 2016	Medical devices - Quality		
			management systems - Requirements for regulatory		
			purposes		
	4	ISO 9001: 2015	Quality management systems – requirements		
	5	ISO 14971: 2019	Medical devices - application of risk management to medical devices		
	6	EN 455-1: 2020	Requirements and testing for freedom from holes		
	7	EN 455-2: 2015	Requirements and testing for physical properties		
	8	EN 455-3: 2015	Requirements and testing for biological evaluation		



Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

		T
9	EN 455-4: 2009	Requirements and testing for
		shelf-life determination
10	ISO 20417:2021	Medical devices - Information to be
		supplied by the manufacturer
11	ISO 15223-1: 2021	Medical devices — Symbols to be
		used with information to be
		supplied by the manufacturer —
		Part 1: General requirements
12	EN ISO 374-1: 2016+A1:	Protective gloves against
	2018	dangerous chemicals and micro-
		organisms - Part 1: Terminology
		and performance requirements for
		chemical risks
13	EN ISO 374-2: 2019	Protective gloves against
		dangerous chemicals and micro-
		organisms - Part 2: Determination
		of resistance to penetration
14	EN ISO 374-4: 2019	Protective gloves against chemicals
		and micro-organisms - Part 4:
		Determination of resistance to
		degradation by chemicals
15	EN ISO 374-5: 2016	Protective gloves against
		dangerous chemicals and micro-
		organisms - Part 5: Terminology
		and performance requirements for
		micro-organisms risks
16	EN 16523-1: 2015+A1:	Determination of material
	2018	resistance to permeation by
		chemicals - Part 1: Permeation by
		liquid chemical under conditions of
		continuous contact
17	EN ISO 21420:2020	Protective gloves — General
		requirements and test methods
		requirements and test methods

We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:

- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentation is retained under the premise of the manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System
- Is following the EU-Type Examination with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.



Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy – Istanbul Turkey Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

Is in conformity to type based on quality assurance of the production process under surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.

• This EU Declaration of Conformity is prepared in accordance with Annex IV of Medical Device Regulation (EU) 2017/745.

Authorized Signatory:

: MURAT YILDIZ Approver

Title : General Manager/CEO

MEDIKAL JAN NIM SIRKETI
Omerli Mah. Separaf Sükrü Koraltı Cad
No:33 4 Al-VAR GylstaNBUL
Büyük Akmice V.D.626 040 4605
el:0213 435/20 64 Fax:0212 438 20 65
: 04 Feb 2024 Signature

Approval Date

Place of Approval : Istanbul, Turkey

