

**EU DECLARATION OF CONFORMITY**

DOC No.	DOC-MYMEDIKAL-ITC-001		
EC Certificate	Not applicable (Self- declared)		
Manufacturer	MY TICARET VE MEDIKAL A.S.		
Manufacturer Address	Ömerli mah General Şükrü Koraltı Cd no:33, 34555 Arnavutkoy/Istanbul, Turkey		
Single Registration Number (SRN)	TR-MF-000018372		
Brand	Mumu Care		
Product Description	Nitrile Powder Free Examination and Protective Gloves		
Intended Purpose	A patient examination glove is a medical device intended for a medical purpose that is worn on the examiner’s hand or finger to prevent contamination between the patient and examiner. Examination glove is intended for medical activities except for surgery.		
Basic UDI-DI	868302002NPVQ		
Size	XS, S, M, L, XL		
European Medical Device Nomenclature (EMDN)	T01020204 (Examination / Treatment Gloves, Nitrile)		
Global Medical Device Nomenclature (GMDN)	56286 (Nitrile Examination/Treatment glove, non-powdered, non-sterile)		
Product Catalogue/Reference Number	MCN01-XS, MCN02-S, MCN03-M, MCN04-L, MCN05-XL		
Product Group Reference Number	SNBE20013-XS, SNBE20014-S, SNBE20015-M, SNBE20016-L, SNBE20017-XL		
Conformity Assessment Route (MDR):	Annex II and Annex III according to EU 2017/745		
Classification & Rule (MDR)	Class I, Rule 1 & Rule 5 according to Annex VIII		
Device Classification (PPER)	Category III		
EU Type-Examination Certificate (PPER)	2777/14815-03/E63-02		
Notified Body (PPER)	EU-Type Examination and Ongoing Conformity by Notified Body SATRA TECHNOLOGY EUROPE LTD Bracetown Business Park, Clonee, D15YN2P, Ireland [CE 2777]		
Applicable Standards			
	No.	Regulation/ Standard Number	Regulation/ Standard Name
	1	MDR (EU) 2017/745	Medical Device Regulation

	2	PPE (EU) 2016/425	Personal Protective Equipment Regulation
	3	ISO 13485: 2016	Medical devices - Quality management systems - Requirements for regulatory purposes
	4	ISO 9001: 2015	Quality management systems – requirements
	5	ISO 14971: 2019	Medical devices - application of risk management to medical devices
	6	EN 455-1: 2020	Requirements and testing for freedom from holes
	7	EN 455-2: 2015	Requirements and testing for physical properties
	8	EN 455-3: 2015	Requirements and testing for biological evaluation
	9	EN 455-4: 2009	Requirements and testing for shelf-life determination
	10	ISO 10993-1: 2018	Biological evaluation of medical devices –Part 1: Evaluation and testing within a risk management process
	11	ISO 10993-10: 2010	Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization
	12	ISO 10993-11: 2017	Biological evaluation of medical devices — Part 11: Tests for systemic toxicity
	13	ISO 20417:2021	Medical devices - Information to be supplied by the manufacturer
	14	ISO 15223-1: 2021	Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements
	15	EN ISO 374-1: 2016+A1: 2018	Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks
	16	EN ISO 374-2: 2019	Protective gloves against dangerous chemicals and micro-organisms - Part 2: Determination of resistance to penetration
	17	EN ISO 374-4: 2019	Protective gloves against chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals

	18	EN ISO 374-5: 2016	Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risks
	19	EN 16523-1: 2015+A1: 2018	Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact
	20	ASTM D 6978-05:2019	Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs
	21	ASTMF1671/F1671-13	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens Using Phi-X174 Bacteriophage Penetration as a Test System


We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:

- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentation is retained under the premise of the manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System.
- Is following the EU-Type Examination with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.
- Is in conformity to type based on quality assurance of the production process under the surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.
- This EU Declaration of Conformity is prepared in accordance with Annex IV of Medical Device Regulation (EU) 2017/745.

**Authorized Signatory:**

Approver : MURAT YILDIZ

Title : General Manager/CEO

Signature :   
MY TICARET VE MEDİKAL ANONİM ŞİRKETİ  
Ömerli Mah. General Şükrü Koraltı Cad  
No:33 Arnavutköy/İSTANBUL  
Büyükdere V.D.626 040 4605  
Tel:0212 438 20 64 Fax:0212 438 20 65  
www.mymedikal.com

Approval Date : 05 Nov 2024

Place of Approval : Istanbul, Turkey

