

EU DECLARATION OF CONFORMITY

DOC No.	DOC-MYMEDİKAL-SRT-003		
EC Certificate	Not applicable (Self- declared)		
Manufacturer	MY TICARET VE MEDİKAL A.S.		
Manufacturer Address	Ömerli mah General Şükrü Koraltı Cd No:33, 34555 Arnavutkoy/Istanbul, Turkey		
Single Registration Number (SRN)	TR-MF-000018372		
Brand	B-good		
Product Description	Latex Powderfree Examination Gloves		
Intended Purpose	A patient examination glove is a medical device intended for a medical purpose that is worn on the examiner’s hand or finger to prevent contamination between the patient and examiner. Examination glove is intended for medical activities except for surgery.		
Basic UDI-DI	868227994LPFX3		
Size	XS, S, M, L, XL		
European Medical Device Nomenclature (EMDN)	T010201 (Examination/Treatment Gloves, Latex)		
Global Medical Device Nomenclature (GMDN)	47172 (Hevea-latex Examination/treatment glove, non-powdered, non-antimicrobial)		
Product Catalogue/Reference Number	BGLP01-XS, BGLP02-S, BGLP03-M, BGLP04-L, BGLP05-XL		
Product Group Reference Number	LO01		
Conformity Assessment Route (MDR):	Annex II and Annex III according to EU 2017/745		
Classification & Rule (MDR)	Class I, Rule 1 & Rule 5		
Device Classification (PPER)	Category III		
EU Type-Examination Certificate (PPER)	2777/10467-05/E15-01		
Notified Body (PPER)	EU-Type Examination and Ongoing Conformity by Notified Body SATRA TECHNOLOGY EUROPE LTD Bracetown Business Park, Clonee, D15YN2P, Ireland [CE 2777]		
Applicable Standards			
	No.	Regulation/ Standard Number	Regulation/ Standard Name
	1	MDR (EU) 2017/745	Medical Device Regulation
	2	PPE (EU) 2016/425	Personal Protective Equipment Regulation
	3	ISO 13485: 2016	Medical devices - Quality management



MY Medikal

MY TICARET VE MEDİKAL A.Ş.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –İstanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.


		systems - Requirements for regulatory purposes
4	ISO 9001: 2015	Quality management systems – requirements
5	ISO 14971: 2019	Medical devices - application of risk management to medical devices
6	EN 455-1: 2020	Requirements and testing for freedom from holes
7	EN 455-2: 2015	Requirements and testing for physical properties
8	EN 455-3: 2015	Requirements and testing for biological evaluation
9	EN 455-4: 2009	Requirements and testing for shelf-life determination
10	ISO 10993-1: 2018	Biological evaluation of medical devices –Part 1: Evaluation and testing within a risk management process
11	ISO 10993-5: 2009	Biological evaluation of medical devices –Part 5: Test for in vitro cytotoxicity
12	ISO 10993-10: 2010	Biological evaluation of medical devices –Part 10: Test for irritation and skin sensitization
13	ASTM F1671: 2013	Standard test method for resistance of materials used in protective clothing to penetration by blood-borne pathogens using phi-x174 bacteriophage penetration as a test system
14	ASTM D3578: 2019	Standard specification for rubber examination gloves
15	ISO 20417:2021	Medical devices - Information to be supplied by the manufacturer
16	ISO 15223-1: 2021	ISO 15223-1 Symbols to be used with information to be supplied by the manufacturer
17	ASTM D7160: 2016	Determination of expiration dating for medical gloves
18	ASTM D7161: 2016	Determination of real time expiration dating of mature medical gloves stored under typical warehouse conditions
19	EN ISO 374-1: 2016+A1: 2018	Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks

	20	EN ISO 374-2: 2019	Protective gloves against dangerous chemicals and micro-organisms - Part 2: Determination of resistance to penetration
	21	EN ISO 374-4: 2019	Protective gloves against chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals
	22	EN ISO 374-5: 2016	Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risks
	23	EN 16523-1: 2015+A1: 2018	Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact
	24	EN ISO 21420: 2020	Protective gloves - General requirements and test methods

We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:

- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentations are retained under the premise of manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System.
- Is following the EU-Type Examination with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.
- Is in conformity to type based on quality assurance of the production process under the surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.
- This EU Declaration of Conformity is prepared in accordance with Annex IV of Medical Device Regulation (EU) 2017/745.

Authorized Signatory:

Approver : MURAT YILDIZ
Title : General Manager/CEO
Signature : 
Approval Date : 15 Feb 2024
Place of Approval : Istanbul, Turkey

MY TICARET VE
MEDİKAL ANONİM ŞİRKETİ
Ömerli Mah. General Şükrü Koraltı Cad
: No:33 Arnavutköy/İSTANBUL
: Büyükdere V.D.626 040 4605
Tel:0212 438 20 64 Fax:0212 438 20 65
www.mymedikal.com

